

Declaration of Conformity

Manufacturer: CMA Microdialysis
Address: 84 October Hill Road, Holliston, Massachusetts 01746-1388, USA
Manufacturing Location: Harvard Bioscience Inc¹. 84 October Hill Road, Holliston, Massachusetts 01746-1388, USA

We herewith declare that the following products:

Product Name: Dual Syringe Pump
Models: CMA 402

Are in conformity of the following applicable European regulations and directives:

2014/35/EU	Low Voltage directive (LVD)
2014/30/EU	Electromagnetic Compatibility directive (EMC)
2011/65/EU	Restriction on the use of certain hazardous substances directive (RoHS) ²
2012/19/EU	Waste electrical and electronic equipment directive (WEEE) ²

Standards used to demonstrate conformity include:

EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use, General Requirements
FCC 47CFR 15B:2014, Class B	
EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use –EMC requirements ⁴
EN ISO 12100:2010	Safety of machinery-General principles for design, risk assessment and risk reduction ³

¹Harvard Bioscience hold an ISO 9001:2008 certificate BSI FM 629138 for its Quality Management System

²Classified as belonging to equipment categories 8 or 9

³ General-purpose Class A low risk device conforming to Annex I & III

⁴Classified and tested as class B equipment in accordance with CISPR 11 definition in a basic electromagnetic environment. This equipment has also been tested and found to comply with the limits for a class A digital device, pursuant to CFR Title 47 part 15 of the FCC rules.

I, the undersigned, hereby declare that the equipment specified is in conformity with the relevant harmonised Union legislation. Signed for and on behalf of CMA at Holliston, USA



Signed:

A handwritten signature in black ink, appearing to read 'Mark Davis'.

Date:
July 06, 2016

Mark Davis
Harvard Bioscience, Senior Director of
Engineering